1

	Application No.	Applicant(s)
Notice of Allowability	09/648,378	SAMSON ET AL.
	Examiner	Art Unit
	Vy Q. Bui	3731
The MAILING DATE of this communication appears on the cover sheet with the correspondence address All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS. This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.		
1. This communication is responsive to <u>amendment 4/30/2004</u> .		
2. The allowed claim(s) is/are 96-100 and 102.		
3. The drawings filed on 8/25/2000 are accepted by the Examiner.		
 4. Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a)		
Attachment(s) 1. Notice of References Cited (PTO-892) 2. Notice of Draftperson's Patent Drawing Review (PTO-948) 3. Information Disclosure Statements (PTO-1449 or PTO/SB/0 Paper No./Mail Date 4. Examiner's Comment Regarding Requirement for Deposit of Biological Material	6. ☐ Interview Summary Paper No./Mail Dat 08), 7. ☐ Examiner's Amendn	e

Application/Control Number: 09/648,378 Page 2

Art Unit: 3731

DETAILED ACTION

Examiner's Amendment

1. An examiner's amendment to the record appears below. Should the changes and/or

additions be unacceptable to applicant, an amendment may be filed as provided by 37

CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later

than the payment of the issue.

2. The application has been amended as follows:

CLAIM 100: line 11: after "said", insert -- catheter --.

Reasons For Allowance

The following is an examiner's statement of reasons for allowance: the last amendment

has clearly distinguished the present invention to the prior art reference. The prior art of

reference has failed to disclose or suggest a catheter comprising an elongate tubular member

defining an inner lumen extending from a proximal end to a distal end, said elongate body

comprising super-elastic alloy braided ribbons sandwiched between a polymeric inner lining

member and an outer covering member, and a first section having a first diameter different from

a second diameter of a second section or the elongate tubular member having a bend diameter

no more than 3.0mm and the bend diameter defining a ratio of a major diameter over a minor

diameter being no more than 1.5.

Claims 96-100 and 102 are allowed.

Art Unit: 3731

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance".

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Vy Q. Bui whose telephone number is 703-306-3420. The examiner can normally be reached on Monday-Thursday.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, McDermott or Shaver can be reached on 703-308-0858. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Vy Q. Bui

17/18/2004

Primary Examiner

Art Unit 3731